

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074/12

Lot No. 978

2024-10

2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO			
B3591	CALIBRATORE a-TPO	978	2024-10	2x6
C3591	CONTROL SERUM a-TPO	978	2024-10	1
		978	2024-10	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	390	≥ 100
Control Serum (IU/mL)	240,7	158,8-636,4

CALIBRATOR

3	0	1	9	7	8	6	5	3	0	4	4	2	8	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



CONTROL SERUM

3	0	1	9	7	8	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 23.6.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.