

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code 81078

Lot No. 972

2024-09

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2471	VARICELLA IgG	972	2024-09	6x6
B2472	CALIBRATORE VARICELLA IgG	972	2024-09	1
C2471	CONTROL SERUM VARICELLA IgG	972	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	≥ 100
Control Serum (Index)	2,6	1,9-3,4
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR

0	1	4	9	7	2	6	0	0	0	0	0	3	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Orsini*




CONTROL SERUM

0	1	4	9	7	2	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Orsini*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Orsini*

DATE 10-10-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.