

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy)

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000- C1- PENICILLOYL G

Code	88000- C1		
Lot No.	963	2024-08	2023-10

REAGENTS

Label	Description	Lot.	Exp	No
D4881	88000- C1 PENICILLOYL G	963	2024-08	2X6
B4880	CALIBRATORE 88000- C1 PENICILLOYL G	963	2024-10	1
C4880	CONTROL SERUM 88000- C1 PENICILLOYL G	963	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	242	> 100
Control Serum (KU/L)	12,1	7,5-15,7


CALIBRATOR

6	5	8	9	6	3	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

6	5	8	9	6	3	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution as Research Use Only (RUO) product.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 14.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.