

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 972



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	972	2024-09	6X6
B3141	CALIBRATORE ANA-Screen	972	2024-09	1
C3141	CONTROL SERUM ANA-Screen	972	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	330	≥ 100
Control Serum (Ndx)	4,0	2,6-4,9

CALIBRATOR


1	1	5	9	7	2	6	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM

1	1	5	9	7	2	6	2	6	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *HB*

DATE 23/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent,