

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**DIESSE**  
**DIESSE**

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHIKUNGUNYA IgG -CHORUS

|         |         |
|---------|---------|
| Code    | 81133   |
| Lot No. | 970     |
|         | 2024-09 |
|         | 2023-10 |

REAGENTS

| Label | Description                 | Lot. | Exp.    | NO  |
|-------|-----------------------------|------|---------|-----|
| D0200 | CHIKUNGUNYA IgG             | 970  | 2024-09 | 6X6 |
| B0200 | CALIBRATORE CHIKUNGUNYA IgG | 970  | 2024-10 | 1   |
| C0200 | CONTROL CHIKUNGUNYA IgG     | 970  | 2024-10 | 1   |

CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                  |     |         |
|------------------|-----|---------|
| Calibrator (mOD) | 523 | ≥ 100   |
| Control (Index)  | 2,8 | 2,1-4,0 |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 0 | 9 | 9 | 7 | 0 | 6 | 0 | 0 | 0 | 0 | 0 | 3 | 3 | 0 | 0 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *RS*



CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 9 | 0 | 9 | 9 | 7 | 0 | 6 | 2 | 1 | 4 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *RS*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Blus*

DATE 25/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.