

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code 86032

Lot No. 978



2024-10



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G			
B3281	CALIBRATORE dsDNA-G	978	2024-10	6X6
C3281	CONTROL SERUM dsDNA-G	978	2024-10	1
		978	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	792	≥ 100
Control Serum (IU/mL)	43,3	32,5-68,2

CALIBRATOR

2 5 1 9 7 8 6 7 2 0 3 1 5 4 3 4 0 0

Label checked with barcode reader

YES ☒

Operator Signature



CONTROL SERUM

2 5 1 9 7 8 6 1 3 2 4 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 26.6.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.