

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME ZIKA IgG -CHORUS

Code 81158

Lot No. 970

2024-09

2023-10

### REAGENTS

Label	Description	Lot.	Exp	No
D4130	Zika IgG	970	2024-09	6X6
B4130	Calibratore Zika IgG	970	2024-10	1
C4130	Control Zika IgG	970	2024-10	1

### CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	472	> 100
Control (AU/ml)	43,2	23,8-62,5


### CALIBRATOR

9	8	3	9	7	0	6	1	2	2	3	1	4	1	4	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



### CONTROL SERUM

9	8	3	9	7	0	6	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES

Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 26-10-2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.