

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR VEDOLIZUMAB

Code 86710

Lot No. 943

2024-06

2023-06

REAGENTS

Label	Description	Lot.	Exp.	NO
D4480	PROMONITOR VEDOLIZUMAB	943	2024-06	6x6
B4480	CALIBRATORE VEDOLIZUMAB	943	2024-08	1
C4480	CONTROL SERUM VEDOLIZUMAB	943	2024-08	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	444	≥ 100
Control Serum (ug/ml)	6,9	6,4-9,0

CALIBRATOR

2	7	8	9	4	3	6	6	1	5	3	5	2	2	9	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RS



CONTROL SERUM

2	7	8	9	4	3	6	3	6	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

RS



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 27/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.