

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code	86601		
Lot No.	971	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9970	CALPROTECTIN	971	2024-09	2X6
B9970	CALIBRATORE CALPROTECTIN	971	2025-01	1
C9970	CONTROL CALPROTECTIN	971	2025-01	1
G9502	SAMPLE DILUENT	971	2025-09	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	417	≥ 100
Control Serum (mg/Kg)	474,0	284,4-585,7

CALIBRATOR

9	2	9	9	7	1	6	2	2	6	5	4	2	6	6	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	2	9	9	7	1	6	2	5	4	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 26/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.