

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code 86060

Lot No. 956-A



2024-07



2023-07

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	956	2024-07	6X6
B3451	CALIBRATORE tTg-G	956-1	2024-07	1
C3451	CONTROL SERUM tTg-G	956-1	2024-07	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	465	≥ 100
Control Serum (AU/mL)	47,9	34,6-68,9

CALIBRATOR

3	4	0	9	5	6	6	1	1	2	2	5	4	9	3	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



CONTROL SERUM

3	4	0	9	5	6	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *MB*

DATE 30.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive