

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code 81060

Lot No. 979



2024-10



2023-10

## REAGENTS

| Label | Description                           | Lot. | Exp.    | NO  |
|-------|---------------------------------------|------|---------|-----|
| D2391 | Helicobacter Pylori IgG               | 979  | 2024-10 | 6X6 |
| B2392 | CALIBRATORE Helicobacter Pylori IgG   | 979  | 2024-10 | 1   |
| C2391 | CONTROL SERUM Helicobacter Pylori IgG | 979  | 2024-10 | 1   |

## CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |      |            |
|--------------------------|------|------------|
| Calibrator (mOD)         | 350  | $\geq 100$ |
| Control Serum (UA/mL)    | 76,6 | 55,3-141,7 |
| Negative Control (UA/mL) | <5   | <5         |

## CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 0 | 8 | 9 | 7 | 9 | 6 | 1 | 2 | 3 | 4 | 2 | 2 | 8 | 5 | 4 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *MB*



## CONTROL SERUM


|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 0 | 8 | 9 | 7 | 9 | 6 | 2 | 5 | 4 | 7 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES



Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 31/10/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive