

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI Ag - CHORUS

Code 81063

Lot No. 971



2024-09



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D6960	Helicobacter pylori Ag	971	2024-09	2X6
B6960	CALIBRATORE Helicobacter pylori Ag	971	2024-12	2
C6960	CONTROL SERUM Helicobacter pylori Ag	971	2024-12	3
G6960	H. pylori Ag Stool Diluent/Negative Control	D0726600-1	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	413	≥ 100
Control + (Index)	2,2	1,9-3,5
Stool Diluent/Negative Control (Index)	0,1	$< 0,5$

CALIBRATOR


9	3	2	9	7	1	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature 



CONTROL SERUM

9	3	2	9	7	1	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader


YES



Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents 

DATE 31.10.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.