

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN - CHORUS

Code 86600

Lot No. 974

2024-09

2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D9500	CALPROTECTIN	974	2024-09	2 X 6
B9500	CALIBRATORE CALPROTECTIN	974	2025-01	1
C9500	CONTROL CALPROTECTIN	974	2025-01	1
G9502	SAMPLE DILUENT	974	2025-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	470	≥ 100
Control (mg/kg)	155,4	92,9-222,7

CALIBRATOR

9	0	0	9	7	4	6	2	1	7	2	2	4	4	3	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *sc*



CONTROL SERUM

9	0	0	9	7	4	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *sc*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *sc*

DATE 03.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.