

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RUBELLA IgG - CHORUS

Code 81030

Lot No. 978

2024-10

2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D2231	RUBELLA IgG	978	2024-10	6X6
B2232	CALIBRATORE RUBELLA IgG	978	2024-10	1
C2231	CONTROL SERUM RUBELLA IgG	978	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	550	≥ 100
Control Serum (IU/mL)	17,5	14,0-27,5
Negative Control (IU/mL)	$< 5,0$	< 5

CALIBRATOR

0	8	1	9	7	8	6	7	2	5	3	5	5	1	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

0	8	1	9	7	8	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

FB

DATE 06/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.