

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074/12

Lot No. 979



2024-10



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	979	2024-10	2x6
B3591	CALIBRATORE a-TPO	979	2024-10	1
C3591	CONTROL SERUM a-TPO	979	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (IU/mL)	405,3	231,6-666,7

CALIBRATOR

3	0	1	9	7	9	6	6	2	3	4	2	2	8	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM


3	0	1	9	7	9	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 06-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.