

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code 81078

Lot No. 973

2024-09

2023-09

REAGENTS

Label	Description	Lot	Exp.	NO
D2471	VARICELLA IgG	973	2024-09	6x6
B2472	CALIBRATORE VARICELLA IgG	973	2024-09	1
C2471	CONTROL SERUM VARICELLA IgG	973	2024-09	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	485	≥ 100
Control Serum (Index)	2,3	1,9-3,4
Negative Control (Index)	0,0	< 0,7

CALIBRATOR

0	1	4	9	7	3	6	0	0	0	0	0	3	8	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *SB*



CONTROL SERUM

0	1	4	9	7	3	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 07.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.