

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME **HELICOBACTER PYLORI IgA-CHORUS**

Code **81062**

Lot No. **965**



2024-08



2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	965	2024-08	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	965	2024-08	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	965	2024-08	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	403	≥ 100
Control Serum (INDEX)	3,2	2,4-4,4
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	2	7	9	6	5	6	0	0	0	0	0	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

0	2	7	9	6	5	6	2	4	4	4	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *Done*

DATE *15/11/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.