

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code 86074

Lot No. 981



2024-10



2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	981	2024-10	6X6
B3591	CALIBRATORE a-TPO	981	2024-10	1
C3591	CONTROL SERUM a-TPO	981	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	391	≥ 100
Control Serum (IU/mL)	350,1	264,3-593,8

CALIBRATOR

3 0 1 9 8 1 6 5 2 3 5 1 2 4 6 7 0 0


Label checked with barcode reader

YES



Operator Signature

SS



CONTROL SERUM

3 0 1 9 8 1 6 2 9 5 4 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature

SS



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

SS

DATE

16-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.