

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 2 IgG RECOMBINANT - CHORUS

Code	81029		
Lot No.	973	2024-09	2023-09

REAGENTS

Label	Description	Lot.	Exp.	NO
D8690	Herpes simplex 2 IgG Recombinant	973	2024-09	6X6
B8690	CALIBRATORE Herpes simplex 2 IgG Recombinant	973	2024-09	1
C8690	CONTROL SERUM Herpes simplex 2 IgG Recombinant	973	2024-09	1


CONTROLS

	Found results	Expected results
Calibrator (mOD)	661	≥ 100
Control Serum (INDEX)	1,7	1,3-2,3
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

2	6	0	9	7	3	6	0	0	0	0	0	5	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	6	0	9	7	3	6	1	3	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 16-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.