

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code 81072

Lot No. 987

2025-02

2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2431	MEASLES IgG	987	2025-02	6X6
B2432	CALIBRATORE MEASLES IgG	987	2025-02	1
C2431	CONTROL SERUM MEASLES IgG	987	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	≥ 100
Control Serum (Index)	3,5	2,2-4,1
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR


0 0 9 9 8 7 7 0 0 0 0 0 3 2 0 0 0 0

Label checked with barcode reader

YES



Operator Signature MB



CONTROL SERUM


0 0 9 9 8 7 7 2 2 4 1 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature MB



**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 23-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.