

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ANTI-ADALIMUMAB

Code	86703		
Lot No.	989	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp	No
D4340	PROMONITOR ANTI-ADALIMUMAB	989	2024-11	6X6
B4340	CALIBRATOR ANTI-ADALIMUMAB	989	2025-06	1
C4340	CONTROL ANTI-ADALIMUMAB	989	2025-06	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	625	> 100
Control (AU/ml)	30,4	27,8-53,8

CALIBRATOR

2	7	7	9	8	9	6	5	2	3	4	0	3	9	6	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	7	7	9	8	9	6	2	1	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 27-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.