

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000- C2- PENICILLOYL V

Code	88000- C2
Lot No.	985
	2024-10
	2023-12

REAGENTS

Label	Description	Lot	Exp	No
D4931	88000- C2 PENICILLOYV	985	2024-10	2X6
B4930	CALIBRATORE 88000- C2 PENICILLOYL V	985	2024-10	1
C4930	CONTROL SERUM 88000- C2 PENICILLOYL V	985	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	> 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	6	2	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature




CONTROL SERUM

6	6	2	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution as Research Use Only (RUO) product.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 06-12-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.