

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ENZYWELL VARICELLA ZOSTER IgG

Code 91078

LOT.NO 136-A



2024-06



2023-03

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93230	006910	MICROPLATE STRIPS	139	2024-11	48 det.
PF93563	007080	CONJUGATE	131	2024-06	1
PF93910	1036	NEGATIVE CONTROL	175-2	2024-06	1
PF92085	006880	POSITIVE CONTROL	119-1	2024-09	1
PF91885	006870	CUT OFF CONTROL	121-1	2024-09	1
PF93619	1034	SUBSTRATE-HS	375-1	2024-09	1
PF93611	009660	DILUENT 2	228	2024-07	1
PF93603	1035	WASHING BUFFER 10x	457-1	2024-08	1
PF93602	1032/RISK0	STOP SOLUTION	215-7	2024-08	1

CONTROLS

	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL	0.021	$\leq 0,6$ Cut Off
CUT OFF CONTROL	0.451	≥ 0.200
POSITIVE CONTROL	1.798	$\geq 1,5$ Cut Off
CV %	4%	$\leq 15\%$

Evaluation of the method according to the package inset in the attached document
This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 06.07.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

IO 09 - 310-V



CODE 91078

CODE 91088

QUALITY CONTROL CERTIFICATE

Annex: Method Evaluation

NAME KIT	VARIGELLA IgG
LOT.	136
DATE OF TEST	06/04/2023

METHOD: ACCORDING TO THE PACKAGE INSERT

1. SOLID PHASE

A) UNIFORMITY BETWEEN PLATES

1st PLATE

2nd PLATE

B) UNIFORMITY WITHIN PLATE

2. BULK CONTROL date: 31/03/2023

- A) Average OD of Cut Off Control
- B) Index of Negative and Positive Sera
- C) Sensitivity with working calibrators

3. FINAL CONTROL date: 06/04/2023

- A) Blank
- B) Average OD of Negative Control
- C) Average OD of Cut Off Control
- D) Average OD of Positive Control

E) HOMOGENEITY OF CUT OFF

F) PRECISION OF FILLING PROCESS

FOUND

OD450 nm 2.098
CV% 4%
OD 450nm 1.908
CV% 4%

OD 450 nm 2.049
CV% 5%

OD 450 nm 0.420
accepted
accepted

OD 450 nm 0.046
OD 450 nm 0.024
OD 450 nm 0.418
OD 450 nm 1.747

No. 6
OD 450 nm 0.464
CV% 3%

EXPECTED

<20%

<20%

< 15%

≥ 0.200

According to specification

According to specification

≤ 0.100

≤ 0,6 Cut Off

≥ 0.200

≥ 1,5 Cut Off

≤ 15%



SATISFACTORY



UNSATISFACTORY

FINAL RESULT: ACCEPTED

Date: 06/04/2023

Signed:
Released: Q.C. Manager