

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgM - CHORUS

Code 81035

Lot No. 990



2025-02



2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2261	MYCOPLASMA PNEUMONIAE IgM			
B2261	CALIBRATORE MYCOPLASMA PNEUMONIAE IgM	990	2025-02	6x6
C2261	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgM	990	2025-02	1
		990	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	580	≥ 100
Control Serum (INDEX)	2,3	1,6-3,0
Negative Control (INDEX)	0,1	$\leq 0,7$

CALIBRATOR

0 6 7 9 9 0 7 0 0 0 0 0 4 3 0 0 0 0

Label checked with barcode reader

YES



Operator Signature

RB



CONTROL SERUM

0 6 7 9 9 0 7 1 6 3 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature

RB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 11.12.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.