

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS CAMPYLOBACTER CONTROL SERUM

Code 85895

LOT.NO

055



2024-11



2023-11

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF90895	B3040	CAMPYLOBACTER CONTROL SERUM	055	2024-11	1

CONTROLS

TESTED WITH:	FOUND RESULTS	EXPECTED RESULTS
CHORUS CAMPYLOBACTER LOT 972	1/64	1/64± 1 Dil.

Evaluation of the method according to the package insert in the attached document
This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED BY:

RESPONSIBLE HEAD, Quality Control Reagents

DATE: 18.12.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.