

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy)

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-C206 - CEPHALOSPORIN

Code	88000- C206		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp	No
D5811	88000-C206 - CEPHALOSPORIN	985	2024-10	2X6
B5810	CALIBRATORE 88000-C206 - CEPHALOSPORIN	985	2025-02	1
C5810	CONTROL SERUM 88000-C206 - CEPHALOSPORIN	985	2025-02	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	314	> 100
Control Serum (KU/L)	8,2	5,9-16,1

CALIBRATOR


6	8	1	9	8	5	6	5	2	0	5	1	2	8	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

6	8	1	9	8	5	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution as Research Use Only (RUO) product.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 13/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.