

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code 86014

Lot No. 975

2024-12

2023-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	975	2024-12	6X6
B3141	CALIBRATORE ANA-Screen	975	2024-12	1
C3141	CONTROL SERUM ANA-Screen	975	2024-12	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	360	$\geq 100$
Control Serum (Ndx)	2,8	2,0-3,6

### CALIBRATOR


1	1	5	9	7	5	7	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HS*



### CONTROL SERUM


1	1	5	9	7	5	7	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *HS*



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

16/12/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.