

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SARS-CoV-2 "Neutralizing" Ab - CHORUS

Code	81408
Lot No.	985
	2024-10
	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D4240	SARS-CoV-2 "Neutralizing" Ab	985	2024-10	6X6
B4240	CALIBRATORE SARS-CoV-2 "Neutralizing" Ab	985	2024-11	1
C4240	CONTROL SERUM SARS-CoV-2 "Neutralizing" Ab	985	2024-11	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	434	≥ 100
Control Serum (BAU/mL)	161,2	100,0-200,0

CALIBRATOR

1	7	0	9	8	5	6	2	1	8	2	4	7	3	2	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature SS



CONTROL SERUM

1	7	0	9	8	5	6	1	0	1	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature SS



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents SS

DATE 15.12.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.