

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GBM - CHORUS

Code	86070		
Lot No.	996	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3551	GBM	996	2024-12	6X6
B3551	CALIBRATORE GBM	996	2024-12	1
C3551	CONTROL SERUM GBM	996	2024-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	558	≥ 100
Control Serum (AU/mL)	46,9	30,0-63,1

CALIBRATOR

2	6	2	9	9	6	6	4	1	9	3	3	3	8	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

2	6	2	9	9	6	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Blue*

DATE 20/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.