

QUALITY CONTROL CERTIFICATE

KIT NAME RESPIRATORY SYNCYTIAL VIRUS IgA - CHORUS

Code 81037

Lot No. 996

2024-09

2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D7671	Respiratory Syncytial Virus IgA	996	2024-09	6X6
B7671	CALIBRATORE Respiratory Syncytial Virus IgA	996	2024-09	1
C7671	CONTROL SERUM Respiratory Syncytial Virus IgA	996	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (Ndx)	3,3	2,1-3,8
Negative Control (Ndx)	0,0	$\leq 0,7$


CALIBRATOR

0 7 8 9 9 6 5 0 0 0 0 0 3 4 0 0 0 0

Label checked with barcode reader

YES

Operator Signature



CONTROL SERUM

0 7 8 9 9 6 5 2 1 3 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES

Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 21/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.