

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME HSV 1 SCREEN - CHORUS

Code 81018

Lot No. 996



2025-03



2023-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2191	HSV 1 SCREEN	996	2025-03	6x6
B2191	CALIBRATORE HSV 1 SCREEN	996	2025-04	1
C2191	CONTROL SERUM HSV 1 SCREEN	996	2025-03	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	520	$\geq 100$
CONTROL SERUM (Index)	3,0	1,3-8,0
NEGATIVE CONTROL (Index)	0,6	$< 0,7$

### CALIBRATOR

0 3 3 9 9 6 7 0 0 0 0 0 5 9 0 0 0 0


Label checked with barcode reader

YES



Operator Signature

FB



### CONTROL SERUM

0 3 3 9 9 6 7 1 3 8 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

FB

DATE 28/12/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.