

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-G - CHORUS

Code 86050

Lot No. 990

2024-11

2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3351	BETA 2-GLYCOPROTEIN-G	990	2024-11	6X6
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	990	2024-11	1
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	990	2024-11	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	408	≥ 100
Control Serum (AU/mL)	26,8	23,8-42,7

CALIBRATOR

2 0 6 9 9 0 6 1 3 4 4 7 4 5 6 0 0 0


Label checked with barcode reader YES ☒ Operator Signature *PB*



CONTROL SERUM

2 0 6 9 9 0 6 1 6 2 9 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature *PB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *03.12.24*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.