

QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgG - CHORUS

Code 81078

Lot No. 997

2024-12

2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D2471	VARICELLA IgG	997	2024-12	6x6
B2472	CALIBRATORE VARICELLA IgG	997	2024-12	1
C2471	CONTROL SERUM VARICELLA IgG	997	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	430	≥ 100
Control Serum (Index)	2,9	2,1-3,9
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR


0 1 4 9 9 7 6 0 0 0 0 0 3 4 0 0 0 0

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


0 1 4 9 9 7 6 2 1 3 9 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 04/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.