

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgG

Code 81261

Lot No. 996



2024-12



2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9640	Chorus Chlamydia trachomatis IgG	996	2024-12	6X6
B9640	CALIBRATOR Chlamydia trachomatis IgG	996	2024-12	1
C9640	POSITIVE CONTROL Chlamydia trachomatis IgG	996	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	395	≥ 100
Positive Control (Index)	1,7	1,2-2,3

CALIBRATOR


2	3	3	9	9	6	6	0	0	0	0	0	5	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

2	3	3	9	9	6	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *AS*

DATE 04/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.