

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

**DIESSE**

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME CHLAMYDOPHILA PNEUMONIAE IgG - CHORUS

Code 81250

Lot No. 988

2024-11

2023-11

### REAGENTS

| Label | Description                         | Lot. | Exp.    | NO  |
|-------|-------------------------------------|------|---------|-----|
| D8590 | Chlamydomphila pneumoniae IgG       | 988  | 2024-11 | 6X6 |
| B8590 | CALIBRATORE Chlamydomphila p. IgG   | 988  | 2024-11 | 1   |
| C8590 | CONTROL SERUM Chlamydomphila p. IgG | 988  | 2024-11 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |         |             |
|--------------------------|---------|-------------|
| Calibrator (mOD)         | 480     | $\geq 100$  |
| Control Serum (AU/mL)    | 67,0    | 30,0-88,8   |
| Negative Control (AU/mL) | $<10,0$ | $\leq 10,0$ |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 2 | 2 | 9 | 8 | 8 | 6 | 7 | 1 | 6 | 2 | 8 | 3 | 2 | 4 | 4 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature MB




### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 2 | 2 | 9 | 8 | 8 | 6 | 2 | 2 | 4 | 1 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader

YES ☒

Operator Signature MB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 12/01/2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.