

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgA-CHORUS

Code	81062		
Lot No.	989	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2401	Helicobacter Pylori IgA	989	2024-11	6X6
B2401	CALIBRATORE Helicobacter Pylori IgA	989	2024-11	1
C2401	CONTROL SERUM Helicobacter Pylori IgA	989	2024-11	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	250	≥ 100
Control Serum (INDEX)	4,7	3,1-5,7
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	2	7	9	8	9	6	0	0	0	0	0	2	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	2	7	9	8	9	6	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 16/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.