

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code	86100		
Lot No.	990	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	990	2024-11	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	990	2024-12	1
C8660	CONTROL SERUM Deamidated Gliadin-G	990	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	629	≥ 100
Control Serum (AU/mL)	57,6	34,1-87,5

CALIBRATOR

2	5	5	9	9	0	6	1	2	1	3	3	4	3	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *Ben*



CONTROL SERUM

2	5	5	9	9	0	6	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *Ben*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 16-01-2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.