

DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-G - CHORUS

Code 86050

Lot No. 993



2024-11



2023-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3351	BETA 2-GLYCOPROTEIN-G	993	2024-11	6X6
B3351	CALIBRATORE BETA 2-GLYCOPROTEIN-G	993	2024-11	1
C3351	CONTROL SERUM BETA 2-GLYCOPROTEIN-G	993	2024-11	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	382	$\geq 100$
Control Serum (AU/mL)	28,7	21,7-41,7

### CALIBRATOR

2	0	6	9	9	3	6	1	3	1	4	4	5	0	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature *[Signature]*



### CONTROL SERUM

2	0	6	9	9	3	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 25.01.2024

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.