

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SS-B - CHORUS

Code 86024

Lot No. 002

2025-01

2024-01

REAGENTS

Label	Description	Lot	Exp.	NO
D3201	SS-B	002	2025-01	6X6
B3201	CALIBRATORE SS-B	002	2025-01	1
C3201	CONTROL SERUM SS-B	002	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	510	≥ 100
Control Serum (AU/mL)	40,0	24,6-55,0

CALIBRATOR

1	9	4	0	0	2	6	3	2	4	3	6	3	8	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*

CONTROL SERUM

1	9	4	0	0	2	6	1	9	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



194002632436385300
SS-B

CALIBRATOR 0.175 mL IVD

LOT 002

2025-01

2°C-8°C

DIESSE Monteriggioni Italy CE



1940026193400000001
SS-B

CONTROL+ 0.425 mL IVD

LOT 002

2025-01

2°C-8°C

DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 26/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.