

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 002

2025-01

2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	002	2025-01	6X6
B3411	CALIBRATORE Gliadin-G	002	2025-01	1
C3411	CONTROL SERUM Gliadin-G	002	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	588	≥ 100
CONTROL SERUM (AU/ml)	37,8	25,9-61,1

CALIBRATOR

5 0 2 0 0 2 6 4 2 6 3 5 4 6 4 3 0 0

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

5 0 2 0 0 2 6 1 5 2 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 06.02.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.