

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR INFLIXIMAB

Code 86700

Lot No. 001 2025-01 2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D4311	PROMONITOR INFLIXIMAB	001	2025-01	6x6
B4312	CALIBRATORE IFX	001	2025-02	1
C4311	CONTROL SERUM IFX	001	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	448	≥ 100
Control Serum (ug/ml)	2,3	1,9-3,3

CALIBRATOR

2 7 4 0 0 1 6 3 2 8 5 2 3 9 7 2 0 0

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2 7 4 0 0 1 6 2 1 3 2 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 13/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.