

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PROMONITOR ADALIMUMAB

Code 86702

Lot No. 001



2025-01



2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D4331	PROMONITOR ADALIMUMAB	001	2025-01	6x6
B4332	CALIBRATORE ADL	001	2025-01	1
C4331	CONTROL SERUM ADL	001	2025-01	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	314	≥ 100
Control Serum ($\mu\text{g/ml}$)	5,3	4,4-7,2

CALIBRATOR

2	7	6	0	0	1	6	3	2	2	3	9	2	5	6	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



CONTROL SERUM

2	7	6	0	0	1	6	3	2	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 19/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.