

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME a-TPO - CHORUS

Code	86074		
Lot No.	002	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3591	a-TPO	002	2025-01	6X6
B3591	CALIBRATORE a-TPO	002	2025-01	1
C3591	CONTROL SERUM a-TPO	002	2025-01	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	340	≥ 100
Control Serum (IU/mL)	337,6	220,0-450,0

CALIBRATOR

3	0	1	0	0	2	6	6	2	5	5	5	2	6	6	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



CONTROL SERUM

3	0	1	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 22/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.