

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code	86032		
Lot No.	010	2025-02	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	010	2025-02	6X6
B3281	CALIBRATORE dsDNA-G	010	2025-02	1
C3281	CONTROL SERUM dsDNA-G	010	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	604	≥ 100
Control Serum (IU/mL)	42,6	29,2-54,5

CALIBRATOR

2	5	1	0	1	0	6	5	2	2	3	3	5	8	3	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HS



CONTROL SERUM

2	5	1	0	1	0	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature HS



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 29.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.