

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code 81034

Lot No. 003



2025-01



2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2251	MYCOPLASMA PNEUMONIAE IgG	003	2025-01	6x6
B2252	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	003	2025-01	1
C2251	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	003	2025-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	495	≥ 100
Control Serum (UA/mL)	18,9	16,7-32,7
Negative Control (UA/mL)	<10,0	<10

CALIBRATOR

0	6	6	0	0	3	6	8	1	3	2	4	6	7	4	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

0	6	6	0	0	3	6	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 04/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.