

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Q FEVER PHASE II IgG - CHORUS

Code 81171

Lot No. 007

2025-02

2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2080	Q FEVER PHASE II IgG	007	2025-02	6x6
B2080	CALIBRATORE Q FEVER PHASE II IgG	007	2025-02	1
C2080	CONTROL SERUM Q FEVER PHASE II IgG	007	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	330	≥ 100
Control Serum (Index)	2,9	1,9-3,5

CALIBRATOR


9 4 4 0 0 7 6 0 0 0 0 0 3 7 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *SB*



CONTROL SERUM


9 4 4 0 0 7 6 1 9 3 5 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *SB*

DATE 04.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.