

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LEGIONELLA PNEUMOPHILA 1 IgG -CHORUS

Code 81091

Lot No. 007



2024-11



2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D7371	LEGIONELLA PNEUMOPHILA 1 IgG	007	2024-11	6X6
B7371	CALIBRATORE L. PNEUMOPHILA 1 IgG	007	2024-11	1
C7371	CONTROL SERUM L. PNEUMOPHILA 1 IgG	007	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	370	≥ 100
Control Serum (INDEX)	3,1	2,0-3,6
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR


1 5 5 0 0 7 5 0 0 0 0 0 3 6 0 0 0 0 0

Label checked with barcode reader

YES



Operator Signature *HB*



CONTROL SERUM

1 5 5 0 0 7 5 2 0 3 6 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.