

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code 81010

Lot No. 997



2024-12



2023-12

REAGENTS

Label	Description	Lot	Exp.	NO
D2161	Cytomegalovirus IgG	997	2024-12	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	997	2024-12	1
C2161	CONTROL SERUM Cytomegalovirus IgG	997	2024-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	300	≥ 100
Control Serum (IU/mL)	12,6	6,8-20,0
Negative Control (IU/mL)	<0,5	<0,5

CALIBRATOR

0	0	1	9	9	7	6	3	2	8	3	9	2	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *MB*



CONTROL SERUM

0	0	1	9	9	7	6	3	2	5	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 13.03.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.