

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

PRODUCT NAME: VACU-CODE

Code	10230		
Lot No.	1124	2025-02	2024-03
Reference IO-09-88			

ON LINE QUALITY CONTROLS (model CDP not enclosed)**QC Passed**

- ☒ sodium citrate solution concentration ☒
- ☒ sodium citrate volume ☒
- ☒ vacuum level (*) ☒
- ☒ correct placement of the cap ☒
- ☒ no extra body inside the product ☒

(*) Only for VACU-CODE and VACU-TEC tubes ☒**PACKAGING QUALITY CONTROL** (model CDP not enclosed)**QC Passed**

- ☒ control of the number of the product for each pack ☒

QUALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)**QC Passed**

- ☒ pack integrity ☒
- ☒ data printed on the product label ☒

QC Passed**N/A****Irradiation treatment Certificate n° date****QC Passed****N/A****Gamma rays treatment process in compliance with the specifications of the supply Agreement**

Document number:

ODV N.A.

***This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.***

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

14.03.2024