

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code 86094

Lot No. 016



2025-03



2024-03

REAGENTS

Label	Description	Lot	Exp.	NO
D8671	ANTI-CCP	016	2025-03	6x6
B8671	CALIBRATORE ANTI-CCP	016	2025-03	1
C8671	CONTROL SERUM ANTI-CCP	016	2025-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	380	≥ 100
Control Serum (AU/mL)	69,7	44,6-92,5

CALIBRATOR


2 5 6 0 1 6 6 4 2 3 4 3 3 3 6 5 0 0

Label checked with barcode reader

YES



Operator Signature *FB*



CONTROL SERUM

2 5 6 0 1 6 6 2 1 4 0 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. *[Signature]*

DATE 19/03/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.